



Note: All Date-Time values are in UTC -05:00

Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D802015630GH	Model: TRIX-8	Battery: OK	Trip #: 62
Description:			

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F after 2 Consecutive
Interval : 5 Minutes	Upper : 158.0 °F after 2 Consecutive
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data

First reading : 1/13/2026 8:22:44 AM	Temperature statistics
Last reading : 2/10/2026 2:57:44 AM	Lowest : 38.5 °F
Elapsed time : 27 Days 18 Hours 40 Minutes	@ 2/9/2026 3:27:44 PM
Total readings : 8000	Highest : 40.1 °F
First evaluated : 2/9/2026 9:27:44 AM	@ 2/9/2026 4:32:44 PM
Last evaluated : 2/10/2026 2:52:44 AM	Average reading: 39.1 °F
Evaluated Time : 17 Hours 30 Minutes	Standard Deviation: 0.3 °F(S)
Evaluated Readings : 210	MKT(Δ H 83.144): 39.1 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00

