

Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D8020156364C	Model: TRIX-8	Battery: OK	Trip #: 66
Description:			

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F after 2 Consecutive
Interval : 5 Minutes	Upper : 158.0 °F after 2 Consecutive
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data


First reading : 1/9/2026 3:00:21 PM	Temperature statistics
Last reading : 2/6/2026 9:35:21 AM	Lowest : -1.7 °F
Elapsed time : 27 Days 18 Hours 40 Minutes	@ 1/30/2026 12:05:21 PM
Total readings : 8000	Highest : 55.6 °F
First evaluated : 1/30/2026 11:55:21 AM	@ 1/30/2026 2:45:21 PM
Last evaluated : 2/6/2026 8:35:21 AM	Average reading: 22.6 °F
Evaluated Time : 6 Days 20 Hours 45 Minutes	Standard Deviation: 7.5 °F(S)
Evaluated Readings : 1977	MKT(Δ H 83.144): 25.2 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Events Info

1/30/2026 2:30:21 PM	Download
2/5/2026 3:15:21 PM	Download