

Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D802028815JH	Model: TRIX-8	Battery: OK	Trip #: 1
Description:			

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data


First reading : 3/6/2026 3:23:42 PM	Temperature statistics
Last reading : 3/16/2026 7:28:42 AM	Lowest : 35.9 °F
Elapsed time : 9 Days 16 Hours 10 Minutes	@ 3/14/2026 8:43:42 AM
Total readings : 2786	Highest : 41.0 °F
First evaluated : 3/13/2026 3:43:42 PM	@ 3/14/2026 2:53:42 PM
Last evaluated : 3/16/2026 7:23:42 AM	Average reading: 36.8 °F
Evaluated Time : 2 Days 15 Hours 45 Minutes	Standard Deviation: 0.7 °F(S)
Evaluated Readings : 765	MKT(Δ H 83.144): 36.8 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Turbo air. - MST-48-N