

## Alarm Status

Lower		Fail
Upper		OK

## Recorder Info

Serial #: D80202881490	Model: TRIX-8	Battery: OK	Trip #: 1
Description:			

## Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

## Recorded Data


First reading : 3/18/2026 1:12:56 PM	Temperature statistics
Last reading : 4/15/2026 7:27:56 AM	Lowest : -13.2 °F
Elapsed time : 27 Days 18 Hours 20 Minutes	@ 4/14/2026 10:12:56 PM
Total readings : 7996	Highest : 77.0 °F
First evaluated : 4/14/2026 1:52:56 PM	@ 4/14/2026 2:37:56 PM
Last evaluated : 4/15/2026 7:22:56 AM	Average reading: -3.8 °F
Evaluated Time : 17 Hours 35 Minutes	Standard Deviation: 19.0 °F(S)
Evaluated Readings : 211	MKT( $\Delta$ H 83.144): 31.3 °F

## Lower Alarm

Triggered: 4/14/2026 2:52:56 PM
Time below/equal: 15 Hours 15 Minutes
Occurrences : 3
°F - Minutes below : 5,155.50

## Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



## Events Info

4/14/2026 2:42:56 PM	Inspection
----------------------	------------

15678396