

Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D80202881730	Model: TRIX-8	Battery: OK	Trip #: 1
Description:			

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 45.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data


First reading : 3/18/2026 1:06:10 PM	Temperature statistics
Last reading : 4/15/2026 7:21:10 AM	Lowest : 6.5 °F
Elapsed time : 27 Days 18 Hours 20 Minutes	@ 4/14/2026 5:01:10 PM
Total readings : 7996	Highest : 17.5 °F
First evaluated : 4/14/2026 3:11:10 PM	@ 4/14/2026 3:11:10 PM
Last evaluated : 4/15/2026 7:16:10 AM	Average reading: 8.4 °F
Evaluated Time : 16 Hours 10 Minutes	Standard Deviation: 2.1 °F(S)
Evaluated Readings : 194	MKT(Δ H 83.144): 8.6 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Contiental -15683708