

Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D802028815JH	Model: TRIX-8	Battery: OK	Trip #: 1
Description:			

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data


First reading : 3/18/2026 1:13:42 PM	Temperature statistics
Last reading : 4/15/2026 7:08:42 AM	Lowest : 0.2 °F
Elapsed time : 27 Days 18 Hours	@ 4/14/2026 4:38:42 PM
Total readings : 7992	Highest : 78.1 °F
First evaluated : 4/14/2026 2:33:42 PM	@ 4/14/2026 2:38:42 PM
Last evaluated : 4/15/2026 7:03:42 AM	Average reading: 4.3 °F
Evaluated Time : 16 Hours 35 Minutes	Standard Deviation: 9.2 °F(S)
Evaluated Readings : 199	MKT(Δ H 83.144): 18.4 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Events Info

4/14/2026 2:38:42 PM	Inspection
----------------------	------------

Contiental Freezer Sn-15683710