

Alarm Status

Lower	✓	OK
Upper	✗	Fail

Recorder Info

Serial #: D802028820ET	Model: TRIX-8	Battery: OK	Trip #: 2
Description: 32598			

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 45.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data


First reading : 3/18/2026 2:00:50 PM	Temperature statistics
Last reading : 4/15/2026 7:55:50 AM	Lowest : 32.5 °F
Elapsed time : 27 Days 18 Hours	@ 4/14/2026 3:15:50 PM
Total readings : 7992	Highest : 75.0 °F
First evaluated : 4/14/2026 2:25:50 PM	@ 4/14/2026 2:30:50 PM
Last evaluated : 4/15/2026 7:50:50 AM	Average reading: 37.1 °F
Evaluated Time : 17 Hours 30 Minutes	Standard Deviation: 5.1 °F(S)
Evaluated Readings : 210	MKT(Δ H 83.144): 39.0 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: 4/14/2026 2:25:50 PM
Time above/equal: 20 Minutes
Occurrences : 1
°F - Minutes above : 472.00



Events Info

4/14/2026 2:40:50 PM	Inspection
----------------------	------------

15684016