

Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D802028816TT	Model: TRIX-8	Battery: OK	Trip #: 2
Description:			

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data

First reading : 5/20/2026 9:19:07 AM	Temperature statistics
Last reading : 5/22/2026 1:34:07 PM	Lowest : 32.7 °F
Elapsed time : 2 Days 4 Hours 20 Minutes	@ 5/21/2026 5:44:07 PM
Total readings : 628	Highest : 37.9 °F
First evaluated : 5/20/2026 10:19:07 AM	@ 5/21/2026 2:49:07 PM
Last evaluated : 5/22/2026 1:34:07 PM	Average reading: 35.4 °F
Evaluated Time : 2 Days 3 Hours 20 Minutes	Standard Deviation: 1.5 °F(S)
Evaluated Readings : 616	MKT(Δ H 83.144): 35.5 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00

