

## Alarm Status

Lower	✓	OK
Upper	✓	OK

## Recorder Info

Serial #: D802028815JH	Model: TRIX-8	Battery: OK	Trip #: 3
Description:			

## Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

## Recorded Data


First reading : 5/21/2026 12:26:02 PM	Temperature statistics
Last reading : 5/25/2026 9:06:02 AM	Lowest : 38.2 °F
Elapsed time : 3 Days 20 Hours 45 Minutes	@ 5/24/2026 2:16:02 PM
Total readings : 1113	Highest : 80.7 °F
First evaluated : 5/21/2026 12:26:02 PM	@ 5/21/2026 12:26:02 PM
Last evaluated : 5/25/2026 9:06:02 AM	Average reading: 40.8 °F
Evaluated Time : 3 Days 20 Hours 45 Minutes	Standard Deviation: 2.5 °F(S)
Evaluated Readings : 1113	MKT( $\Delta$ H 83.144): 41.1 °F

## Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

## Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Source: TRIX-8- D802028815JH Started 5\_21\_2026, Finished 5\_25\_2026.ltd

File Location: U:\My LogTag Data

Created : 5/25/2026 9:08:59 AM