

Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D802028815JH	Model: TRIX-8	Battery: OK	Trip #: 2
Description:			

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data


First reading : 5/19/2026 3:42:32 PM	Temperature statistics
Last reading : 5/21/2026 9:07:32 AM	Lowest : 36.0 °F
Elapsed time : 1 Day 17 Hours 30 Minutes	@ 5/19/2026 5:47:32 PM
Total readings : 498	Highest : 52.7 °F
First evaluated : 5/19/2026 4:42:32 PM	@ 5/20/2026 4:27:32 PM
Last evaluated : 5/21/2026 9:07:32 AM	Average reading: 38.7 °F
Evaluated Time : 1 Day 16 Hours 30 Minutes	Standard Deviation: 2.6 °F(S)
Evaluated Readings : 486	MKT(Δ H 83.144): 39.0 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Source: TRIX-8- D802028815JH Started 5_19_2026, Finished 5_21_2026.ltd

File Location: U:\My LogTag Data

Created : 5/21/2026 9:13:55 AM