

## Alarm Status

Lower	✓	OK
Upper	✓	OK

## Recorder Info

Serial #: D802028816TT	Model: TRIX-8	Battery: OK	Trip #: 4
Description:			

## Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

## Recorded Data


First reading : 5/28/2026 11:07:19 AM	Temperature statistics
Last reading : 5/29/2026 9:07:19 AM	Lowest : 36.2 °F
Elapsed time : 22 Hours 5 Minutes	@ 5/28/2026 12:12:19 PM
Total readings : 265	Highest : 83.9 °F
First evaluated : 5/28/2026 11:07:19 AM	@ 5/28/2026 11:07:19 AM
Last evaluated : 5/29/2026 9:07:19 AM	Average reading: 38.0 °F
Evaluated Time : 22 Hours 5 Minutes	Standard Deviation: 3.3 °F(S)
Evaluated Readings : 265	MKT( $\Delta$ H 83.144): 38.9 °F

## Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

## Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Source: TRIX-8- D802028816TT Started 5\_28\_2026, Finished 5\_29\_2026.ltd

File Location: U:\My LogTag Data

Created : 5/29/2026 9:12:17 AM