

Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D802028821O5	Model: TRIX-8	Battery: OK	Trip #: 1
Description:			

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data

First reading : 5/22/2026 12:26:13 PM	Temperature statistics
Last reading : 5/25/2026 2:36:13 PM	Lowest : 35.8 °F
Elapsed time : 3 Days 2 Hours 15 Minutes	@ 5/23/2026 6:11:13 AM
Total readings : 891	Highest : 79.9 °F
First evaluated : 5/22/2026 12:26:13 PM	@ 5/22/2026 12:26:13 PM
Last evaluated : 5/25/2026 2:36:13 PM	Average reading: 37.1 °F
Evaluated Time : 3 Days 2 Hours 15 Minutes	Standard Deviation: 1.8 °F(S)
Evaluated Readings : 891	MKT(Δ H 83.144): 37.4 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00

