

Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D802028822YH	Model: TRIX-8	Battery: OK	Trip #: 2	Description:
------------------------	---------------	-------------	-----------	--------------

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data

First reading : 6/1/2026 8:49:30 AM	Temperature statistics
Last reading : 6/8/2026 9:54:30 AM	Lowest : 36.1 °F
Elapsed time : 7 Days 1 Hour 10 Minutes	@ 6/3/2026 9:19:30 AM
Total readings : 2030	Highest : 75.2 °F
First evaluated : 6/1/2026 8:49:30 AM	@ 6/1/2026 8:49:30 AM
Last evaluated : 6/8/2026 9:54:30 AM	Average reading: 38.1 °F
Evaluated Time : 7 Days 1 Hour 10 Minutes	Standard Deviation: 1.6 °F(S)
Evaluated Readings : 2030	MKT(ΔH 83.144): 38.2 °F

Lower Alarm


Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00

Temp °F

Temp °F



Source: TRIX-8- D802028822YH Started 6_1_2026, Finished 6_8_2026.ltd

File Location: U:\My LogTag Data

Created : 6/8/2026 9:55:03 AM