

Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D802028816TT	Model: TRIX-8	Battery: OK	Trip #: 9	Description:
------------------------	---------------	-------------	-----------	--------------

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data

First reading : 6/5/2026 9:56:02 AM	Temperature statistics
Last reading : 6/8/2026 8:41:02 AM	Lowest : 27.2 °F
Elapsed time : 2 Days 22 Hours 50 Minutes	@ 6/5/2026 12:46:02 PM
Total readings : 850	Highest : 67.8 °F
First evaluated : 6/5/2026 9:56:02 AM	@ 6/5/2026 9:56:02 AM
Last evaluated : 6/8/2026 8:41:02 AM	Average reading: 30.5 °F
Evaluated Time : 2 Days 22 Hours 50 Minutes	Standard Deviation: 3.4 °F(S)
Evaluated Readings : 850	MKT(Δ H 83.144): 31.1 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00

