

Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D802028819NT	Model: TRIX-8	Battery: OK	Trip #: 4	Description:
------------------------	---------------	-------------	-----------	--------------

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data


First reading : 6/8/2026 3:13:52 PM	Temperature statistics
Last reading : 6/10/2026 9:33:52 AM	Lowest : 34.8 °F
Elapsed time : 1 Day 18 Hours 25 Minutes	@ 6/10/2026 4:38:52 AM
Total readings : 509	Highest : 80.3 °F
First evaluated : 6/8/2026 3:13:52 PM	@ 6/8/2026 3:13:52 PM
Last evaluated : 6/10/2026 9:33:52 AM	Average reading: 36.1 °F
Evaluated Time : 1 Day 18 Hours 25 Minutes	Standard Deviation: 2.5 °F(S)
Evaluated Readings : 509	MKT(Δ H 83.144): 36.6 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Source: TRIX-8- D802028819NT Started 6_8_2026, Finished 6_10_2026.ltd

File Location: U:\My LogTag Data

Created : 6/10/2026 9:35:17 AM