

Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D802028816TT	Model: TRIX-8	Battery: OK	Trip #: 12	Description:
------------------------	---------------	-------------	------------	--------------

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data

First reading : 6/23/2026 11:48:35 AM	Temperature statistics
Last reading : 6/30/2026 12:33:35 PM	Lowest : 37.9 °F
Elapsed time : 7 Days 50 Minutes	@ 6/25/2026 10:13:35 PM
Total readings : 2026	Highest : 80.4 °F
First evaluated : 6/23/2026 11:48:35 AM	@ 6/23/2026 11:48:35 AM
Last evaluated : 6/30/2026 12:33:35 PM	Average reading: 40.7 °F
Evaluated Time : 7 Days 50 Minutes	Standard Deviation: 4.8 °F(S)
Evaluated Readings : 2026	MKT(Δ H 83.144): 41.7 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00

