

Alarm Status

Lower	✓	OK
Upper	✗	Fail

Recorder Info

Serial #: D80202881730	Model: TRIX-8	Battery: OK	Trip #: 9	Description:
------------------------	---------------	-------------	-----------	--------------

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 45.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data

First reading : 6/18/2026 12:21:06 PM	Temperature statistics
Last reading : 6/30/2026 2:31:06 PM	Lowest : 30.7 °F
Elapsed time : 12 Days 2 Hours 15 Minutes	@ 6/23/2026 10:46:06 AM
Total readings : 3483	Highest : 77.6 °F
First evaluated : 6/18/2026 12:21:06 PM	@ 6/19/2026 4:01:06 AM
Last evaluated : 6/30/2026 2:31:06 PM	Average reading: 39.1 °F
Evaluated Time : 12 Days 2 Hours 15 Minutes	Standard Deviation: 9.5 °F(S)
Evaluated Readings : 3483	MKT(ΔH 83.144): 44.1 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: 6/18/2026 12:21:06 PM
Time above/equal: 1 Day 22 Hours 30 Minutes
Occurrences : 94
°F - Minutes above : 32,149.50

