

Alarm Status

Lower	✓	OK
Upper	✗	Fail

Recorder Info

Serial #: D80202881730	Model: TRIX-8	Battery: OK	Trip #: 8	Description:
------------------------	---------------	-------------	-----------	--------------

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 45.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data

First reading : 6/17/2026 12:51:25 PM	Temperature statistics
Last reading : 6/18/2026 12:06:25 PM	Lowest : -2.5 °F
Elapsed time : 23 Hours 20 Minutes	@ 6/17/2026 3:31:25 PM
Total readings : 280	Highest : 76.1 °F
First evaluated : 6/17/2026 12:51:25 PM	@ 6/17/2026 12:51:25 PM
Last evaluated : 6/18/2026 12:06:25 PM	Average reading: 3.5 °F
Evaluated Time : 23 Hours 20 Minutes	Standard Deviation: 6.2 °F(S)
Evaluated Readings : 280	MKT(Δ H 83.144): 10.2 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: 6/17/2026 12:51:25 PM
Time above/equal: 5 Minutes
Occurrences : 1
°F - Minutes above : 155.50

