



Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D802015631QT	Model: TRIX-8	Battery: OK	Trip #: 52	Description:
------------------------	---------------	-------------	------------	--------------

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F after 2 Consecutive
Interval : 5 Minutes	Upper : 158.0 °F after 2 Consecutive
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data


First reading : 6/17/2026 12:51:29 PM	Temperature statistics
Last reading : 6/18/2026 12:16:29 PM	Lowest : 39.0 °F
Elapsed time : 23 Hours 30 Minutes	@ 6/18/2026 3:41:29 AM
Total readings : 282	Highest : 76.6 °F
First evaluated : 6/17/2026 12:51:29 PM	@ 6/17/2026 12:51:29 PM
Last evaluated : 6/18/2026 12:16:29 PM	Average reading: 41.7 °F
Evaluated Time : 23 Hours 30 Minutes	Standard Deviation: 4.0 °F(S)
Evaluated Readings : 282	MKT(ΔH 83.144): 42.5 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Source: TRI-X-8- D802015631QT Started 6_17_2026, Finished 6_18_2026.ltd

File Location: U:\My LogTag Data

Created : 6/18/2026 12:18:18 PM