

Alarm Status

Lower	✓	OK
Upper	✗	Fail

Recorder Info

Serial #: D802028820ET	Model: TRIX-8	Battery: OK	Trip #: 5	Description: 32598
------------------------	---------------	-------------	-----------	--------------------

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 45.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data

First reading : 6/17/2026 12:56:25 PM	Temperature statistics
Last reading : 6/18/2026 10:51:25 AM	Lowest : 33.3 °F
Elapsed time : 22 Hours	@ 6/17/2026 3:01:25 PM
Total readings : 264	Highest : 77.9 °F
First evaluated : 6/17/2026 12:56:25 PM	@ 6/17/2026 12:56:25 PM
Last evaluated : 6/18/2026 10:51:25 AM	Average reading: 37.8 °F
Evaluated Time : 22 Hours	Standard Deviation: 4.2 °F(S)
Evaluated Readings : 264	MKT(Δ H 83.144): 38.9 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: 6/17/2026 12:56:25 PM
Time above/equal: 15 Minutes
Occurrences : 2
°F - Minutes above : 314.50

