



Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D802028822YH	Model: TRIX-8	Battery: OK	Trip #: 11	Description:
------------------------	---------------	-------------	------------	--------------

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data


First reading : 7/3/2026 2:41:43 PM	Temperature statistics
Last reading : 7/6/2026 1:11:43 PM	Lowest : 33.6 °F
Elapsed time : 2 Days 22 Hours 35 Minutes	@ 7/4/2026 9:51:43 AM
Total readings : 847	Highest : 87.3 °F
First evaluated : 7/3/2026 2:41:43 PM	@ 7/3/2026 2:41:43 PM
Last evaluated : 7/6/2026 1:11:43 PM	Average reading: 36.4 °F
Evaluated Time : 2 Days 22 Hours 35 Minutes	Standard Deviation: 2.3 °F(S)
Evaluated Readings : 847	MKT(ΔH 83.144): 36.9 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Source: TRIX-8- D802028822YH Started 7_3_2026, Finished 7_6_2026.ltd

File Location: U:\My LogTag Data

Created : 7/6/2026 1:20:03 PM