



Alarm Status

Lower		Fail
Upper		OK

Recorder Info

Serial #: D802028818DH	Model: TRIX-8	Battery: OK	Trip #: 5	Description:
------------------------	---------------	-------------	-----------	--------------

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data


First reading : 7/10/2026 2:09:59 PM	Temperature statistics
Last reading : 7/14/2026 11:04:59 AM	Lowest : -12.1 °F
Elapsed time : 3 Days 21 Hours	@ 7/13/2026 1:59:59 PM
Total readings : 1116	Highest : 18.4 °F
First evaluated : 7/13/2026 1:09:59 PM	@ 7/13/2026 3:34:59 PM
Last evaluated : 7/14/2026 11:04:59 AM	Average reading: 6.2 °F
Evaluated Time : 22 Hours	Standard Deviation: 5.2 °F(S)
Evaluated Readings : 264	MKT(ΔH 83.144): 7.1 °F

Lower Alarm

Triggered: 7/13/2026 1:09:59 PM
Time below/equal: 1 Hour 40 Minutes
Occurrences : 2
°F - Minutes below : 412.50

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Events Info

7/13/2026 3:34:59 PM	Download
----------------------	----------

Source: TRIX-8- D802028818DH Started 7_10_2026, Finished 7_14_2026.ltd

File Location: U:\My LogTag Data

Created : 7/14/2026 11:09:32 AM