



Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D8020288238O	Model: TRIX-8	Battery: OK	Trip #: 11	Description:
------------------------	---------------	-------------	------------	--------------

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data


First reading : 7/13/2026 3:28:37 PM	Temperature statistics
Last reading : 7/14/2026 11:13:37 AM	Lowest : 49.8 °F
Elapsed time : 19 Hours 50 Minutes	@ 7/14/2026 6:13:37 AM
Total readings : 238	Highest : 52.4 °F
First evaluated : 7/13/2026 4:28:37 PM	@ 7/13/2026 4:28:37 PM
Last evaluated : 7/14/2026 11:13:37 AM	Average reading: 51.0 °F
Evaluated Time : 18 Hours 50 Minutes	Standard Deviation: 0.8 °F(S)
Evaluated Readings : 226	MKT(ΔH 83.144): 51.0 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Source: TRI-X-8- D8020288238O Started 7_13_2026, Finished 7_14_2026.ltd

File Location: U:\My LogTag Data

Created : 7/14/2026 11:19:17 AM