



Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D802028818DH	Model: TRIX-8	Battery: OK	Trip #: 7	Description:
------------------------	---------------	-------------	-----------	--------------

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data


First reading : 7/24/2026 8:33:24 AM	Temperature statistics
Last reading : 7/27/2026 8:23:24 AM	Lowest : 36.9 °F
Elapsed time : 2 Days 23 Hours 55 Minutes	@ 7/25/2026 4:23:24 AM
Total readings : 863	Highest : 66.5 °F
First evaluated : 7/24/2026 3:33:24 PM	@ 7/27/2026 8:23:24 AM
Last evaluated : 7/27/2026 8:23:24 AM	Average reading: 39.2 °F
Evaluated Time : 2 Days 16 Hours 55 Minutes	Standard Deviation: 2.3 °F(S)
Evaluated Readings : 779	MKT(ΔH 83.144): 39.4 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Source: TRI-X-8- D802028818DH Started 7_24_2026, Finished 7_27_2026.ltd

File Location: U:\My LogTag Data

Created : 7/27/2026 8:27:31 AM