



Alarm Status

Lower		Fail
Upper		OK

Recorder Info

Serial #: D802015631QT	Model: TRIX-8	Battery: OK	Trip #: 54	Description:
------------------------	---------------	-------------	------------	--------------

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F after 2 Consecutive
Interval : 5 Minutes	Upper : 158.0 °F after 2 Consecutive
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data


First reading : 7/23/2026 9:26:59 AM	Temperature statistics
Last reading : 7/24/2026 10:26:59 AM	Lowest : -13.6 °F
Elapsed time : 1 Day 1 Hour 5 Minutes	@ 7/24/2026 8:56:59 AM
Total readings : 301	Highest : 11.1 °F
First evaluated : 7/23/2026 10:26:59 AM	@ 7/23/2026 5:31:59 PM
Last evaluated : 7/24/2026 10:26:59 AM	Average reading: -6.8 °F
Evaluated Time : 1 Day 5 Minutes	Standard Deviation: 3.8 °F(S)
Evaluated Readings : 289	MKT(ΔH 83.144): -6.0 °F

Lower Alarm

Triggered: 7/23/2026 10:31:59 AM
Time below/equal: 22 Hours 5 Minutes
Occurrences : 4
°F - Minutes below : 4,880.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Source: TRIX-8- D802015631QT Started 7_23_2026, Finished 7_24_2026.ltd

File Location: U:\My LogTag Data

Created : 7/24/2026 10:31:55 AM