



Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D802028816TT	Model: TRIX-8	Battery: OK	Trip #: 13	Description:
------------------------	---------------	-------------	------------	--------------

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data


First reading : 8/3/2026 11:11:32 AM	Temperature statistics
Last reading : 8/5/2026 9:01:32 AM	Lowest : 31.0 °F
Elapsed time : 1 Day 21 Hours 55 Minutes	@ 8/5/2026 8:41:32 AM
Total readings : 551	Highest : 47.5 °F
First evaluated : 8/4/2026 1:11:32 PM	@ 8/4/2026 8:16:32 PM
Last evaluated : 8/5/2026 9:01:32 AM	Average reading: 36.0 °F
Evaluated Time : 19 Hours 55 Minutes	Standard Deviation: 3.8 °F(S)
Evaluated Readings : 239	MKT(ΔH 83.144): 36.6 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Source: TRI-X-8- D802028816TT Started 8_3_2026, Finished 8_5_2026.ltd

File Location: U:\My LogTag Data

Created : 8/5/2026 9:06:32 AM