



Alarm Status

Lower		Fail
Upper		OK

Recorder Info

Serial #: D8020288238O	Model: TRIX-8	Battery: OK	Trip #: 13	Description:
------------------------	---------------	-------------	------------	--------------

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data


First reading : 8/3/2026 10:29:14 AM	Temperature statistics
Last reading : 8/5/2026 9:19:14 AM	Lowest : -10.7 °F
Elapsed time : 1 Day 22 Hours 55 Minutes	@ 8/5/2026 7:54:14 AM
Total readings : 563	Highest : 81.3 °F
First evaluated : 8/3/2026 10:29:14 AM	@ 8/3/2026 10:29:14 AM
Last evaluated : 8/5/2026 9:19:14 AM	Average reading: -8.2 °F
Evaluated Time : 1 Day 22 Hours 55 Minutes	Standard Deviation: 9.0 °F(S)
Evaluated Readings : 563	MKT(ΔH 83.144): 12.4 °F

Lower Alarm

Triggered: 8/3/2026 12:59:14 PM
Time below/equal: 1 Day 20 Hours 25 Minutes
Occurrences : 1
°F - Minutes below : 15,114.50

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Source: TRI-X-8- D8020288238O Started 8_3_2026, Finished 8_5_2026.ltd

File Location: U:\My LogTag Data

Created : 8/5/2026 9:20:39 AM