



Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D8020288238O	Model: TRIX-8	Battery: OK	Trip #: 14	Description:
------------------------	---------------	-------------	------------	--------------

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data


First reading : 8/5/2026 3:39:35 PM	Temperature statistics
Last reading : 8/7/2026 8:44:35 AM	Lowest : 34.4 °F
Elapsed time : 1 Day 17 Hours 10 Minutes	@ 8/6/2026 3:54:35 PM
Total readings : 494	Highest : 38.2 °F
First evaluated : 8/5/2026 5:39:35 PM	@ 8/6/2026 6:44:35 AM
Last evaluated : 8/7/2026 8:44:35 AM	Average reading: 36.7 °F
Evaluated Time : 1 Day 15 Hours 10 Minutes	Standard Deviation: 0.5 °F(S)
Evaluated Readings : 470	MKT(ΔH 83.144): 36.7 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Source: TRIX-8- D8020288238O Started 8_5_2026, Finished 8_7_2026.ltd

File Location: U:\My LogTag Data

Created : 8/7/2026 8:50:15 AM