



Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D802028815JH	Model: TRIX-8	Battery: OK	Trip #: 17	Description:
------------------------	---------------	-------------	------------	--------------

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data


First reading : 8/5/2026 3:48:08 PM	Temperature statistics
Last reading : 8/7/2026 8:38:08 AM	Lowest : 35.6 °F
Elapsed time : 1 Day 16 Hours 55 Minutes	@ 8/5/2026 5:43:08 PM
Total readings : 491	Highest : 39.2 °F
First evaluated : 8/5/2026 4:48:08 PM	@ 8/6/2026 3:18:08 PM
Last evaluated : 8/7/2026 8:38:08 AM	Average reading: 36.1 °F
Evaluated Time : 1 Day 15 Hours 55 Minutes	Standard Deviation: 0.3 °F(S)
Evaluated Readings : 479	MKT(ΔH 83.144): 36.1 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Source: TRIX-8- D802028815JH Started 8_5_2026, Finished 8_7_2026.ltd

File Location: U:\My LogTag Data

Created : 8/7/2026 8:44:13 AM