



Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D802015631QT	Model: TRIX-8	Battery: OK	Trip #: 59	Description:
------------------------	---------------	-------------	------------	--------------

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F after 2 Consecutive
Interval : 5 Minutes	Upper : 158.0 °F after 2 Consecutive
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data


First reading : 9/10/2026 12:40:38 PM	Temperature statistics
Last reading : 9/16/2026 7:55:38 AM	Lowest : 15.6 °F
Elapsed time : 5 Days 19 Hours 20 Minutes	@ 9/14/2026 5:00:38 PM
Total readings : 1672	Highest : 69.1 °F
First evaluated : 9/14/2026 12:40:38 PM	@ 9/15/2026 8:15:38 AM
Last evaluated : 9/16/2026 7:55:38 AM	Average reading: 35.2 °F
Evaluated Time : 1 Day 19 Hours 20 Minutes	Standard Deviation: 8.6 °F(S)
Evaluated Readings : 520	MKT(ΔH 83.144): 37.4 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Events Info

9/15/2026 8:20:38 AM	Download
----------------------	----------

Source: TRIX-8- D802015631QT Started 9_10_2026, Finished 9_16_2026.ltd

File Location: U:\My LogTag Data

Created : 9/16/2026 8:02:31 AM