



Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D802028815JH	Model: TRIX-8	Battery: OK	Trip #: 23	Description:
------------------------	---------------	-------------	------------	--------------

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data


First reading : 9/10/2026 10:02:45 AM	Temperature statistics
Last reading : 9/11/2026 7:47:45 AM	Lowest : 37.5 °F
Elapsed time : 21 Hours 50 Minutes	@ 9/10/2026 5:17:45 PM
Total readings : 262	Highest : 44.1 °F
First evaluated : 9/10/2026 1:22:45 PM	@ 9/11/2026 1:47:45 AM
Last evaluated : 9/11/2026 7:47:45 AM	Average reading: 38.5 °F
Evaluated Time : 18 Hours 30 Minutes	Standard Deviation: 1.2 °F(S)
Evaluated Readings : 222	MKT(ΔH 83.144): 38.5 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Source: TRI-X-8- D802028815JH Started 9_10_2026, Finished 9_11_2026.ltd

File Location: U:\My LogTag Data

Created : 9/11/2026 7:54:55 AM