



Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D802028819NT	Model: TRIX-8	Battery: OK	Trip #: 15	Description:
------------------------	---------------	-------------	------------	--------------

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data


First reading : 9/24/2026 2:01:14 PM	Temperature statistics
Last reading : 9/28/2026 7:46:14 AM	Lowest : 32.6 °F
Elapsed time : 3 Days 17 Hours 50 Minutes	@ 9/25/2026 11:26:14 PM
Total readings : 1078	Highest : 39.4 °F
First evaluated : 9/25/2026 6:01:14 PM	@ 9/25/2026 6:41:14 PM
Last evaluated : 9/28/2026 7:46:14 AM	Average reading: 35.8 °F
Evaluated Time : 2 Days 13 Hours 50 Minutes	Standard Deviation: 2.1 °F(S)
Evaluated Readings : 742	MKT(ΔH 83.144): 35.9 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Source: TRIX-8- D802028819NT Started 9_24_2026, Finished 9_28_2026.ltd

File Location: U:\My LogTag Data

Created : 9/28/2026 7:52:07 AM