



Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D8020288238O	Model: TRIX-8	Battery: OK	Trip #: 23	Description:
------------------------	---------------	-------------	------------	--------------

Recorder Configuration

Start type : Push button start	Temperature alarms
Start delay : None	Lower : -4.0 °F
Interval : 5 Minutes	Upper : 158.0 °F
Alert indicator : Enabled lower & upper	
OK indicator : Enabled	

Recorded Data


First reading : 10/21/2026 4:58:18 PM	Temperature statistics
Last reading : 10/23/2026 8:08:18 AM	Lowest : 37.8 °F
Elapsed time : 1 Day 15 Hours 15 Minutes	@ 10/22/2026 8:08:18 AM
Total readings : 471	Highest : 74.3 °F
First evaluated : 10/21/2026 4:58:18 PM	@ 10/21/2026 5:03:18 PM
Last evaluated : 10/23/2026 8:08:18 AM	Average reading: 39.8 °F
Evaluated Time : 1 Day 15 Hours 15 Minutes	Standard Deviation: 3.3 °F(S)
Evaluated Readings : 471	MKT(ΔH 83.144): 40.4 °F

Lower Alarm

Triggered: (none)
Time below/equal: (none)
Occurrences : 0
°F - Minutes below : 0.00

Upper Alarm

Triggered: (none)
Time above/equal: (none)
Occurrences : 0
°F - Minutes above : 0.00



Source: TRIx-8- D8020288238O Started 10_21_2026, Finished 10_23_2026.ltd

File Location: U:\My LogTag Data

Created : 10/23/2026 8:12:10 AM