LogTag

Recorder has been Downloaded 11/6/2026 2:45:43 PM

Note: All Date-Time values are in UTC -05:00 DayLight

Alarm Status

Lower	✓	OK
Upper	✓	OK

Recorder Info

Serial #: D802015631QT Model: TRIX-8 Battery: OK Trip #: 62 Description:

Recorder Configuration

Start type :	Push button start	Temperature alarms	
Start delay :	None	Lower : -4.0 °F after 2 Consecutive	
Interval :	5 Minutes	Upper : 158.0 °F after 2 Consecutive	
Alert indicator :	Enabled lower & upper		
OK indicator :	Enabled		

Recorded Data

First reading :	10/30/2026 3:29:10 PM	Temperature statistics	
Last reading :	11/6/2026 3:44:10 PM	Lowest :	29.6 °F
Elapsed time :	7 Days 20 Minutes	@	10/31/2026 10:24:10 AM
Total readings :	2020	Highest :	70.7 °F
First evaluated :	10/30/2026 3:29:10 PM	@	10/30/2026 3:29:10 PM
Last evaluated :	11/6/2026 3:44:10 PM	Average reading:	34.4 °F
Evaluated Time :	7 Days 20 Minutes	Standard Deviation:	3.5 °F(S)
Evaluated Readings :	2020	MKT(ΔH 83.144):	34.9 °F

Lower Alarm

Triggered:	(none)
Time below/equal:	(none)
Occurrences :	0
°F - Minutes below :	0.00


Upper Alarm

Triggered:	(none)
Time above/equal:	(none)
Occurrences :	0
°F - Minutes above :	0.00

Source: TRIX-8- D802015631QT Started 10_30_2026, Finished 11_6_2026.ltd

File Location: U:\My LogTag Data

Created : 11/6/2026 2:53:23 PM



Source: TRIX-8- D802015631QT Started 10_30_2026, Finished 11_6_2026.ltd

File Location: U:\My LogTag Data

Created : 11/6/2026 2:53:23 PM