LogTag

Recorder has been Downloaded 11/4/2026 10:45:14 AM

Note: All Date-Time values are in UTC -05:00 DayLight

Alarm Status

Lower	✗	Fail
Upper	✗	Fail

Recorder Info

Serial #: D802028820ET Model: TRIX-8 Battery: OK Trip #: 21 Description: 32598

Recorder Configuration

Start type :	Push button start	Temperature alarms	
Start delay :	None	Lower :	-4.0 °F
Interval :	5 Minutes	Upper :	45.0 °F
Alert indicator :	Enabled lower & upper		
OK indicator :	Enabled		

Recorded Data


First reading :	11/3/2026 9:11:06 AM	Temperature statistics	
Last reading :	11/4/2026 11:46:06 AM	Lowest :	-26.8 °F
Elapsed time :	1 Day 2 Hours 40 Minutes	@	11/3/2026 9:56:06 AM
Total readings :	320	Highest :	70.1 °F
First evaluated :	11/3/2026 9:11:06 AM	@	11/3/2026 9:11:06 AM
Last evaluated :	11/4/2026 11:46:06 AM	Average reading:	-25.0 °F
Evaluated Time :	1 Day 2 Hours 40 Minutes	Standard Deviation:	5.7 °F(S)
Evaluated Readings :	320	MKT(ΔH 83.144):	-4.8 °F

Lower Alarm

Triggered:	11/3/2026 9:16:06 AM
Time below/equal:	1 Day 2 Hours 30 Minutes
Occurrences :	1
°F - Minutes below :	34,012.50

Upper Alarm

Triggered:	11/3/2026 9:11:06 AM
Time above/equal:	5 Minutes
Occurrences :	1
°F - Minutes above :	125.50



Source: TRIX-8- D802028820ET Started 11_3_2026, Finished 11_4_2026.ltd

File Location: U:\My LogTag Data

Created : 11/4/2026 10:45:49 AM